

Potency Results

Sample Name: *Hemp A - Brr Berry (J)*

Client:

Client Batch ID:

Sample ID: rC-H-185-D2081

Date Sampled: 11/29/2023

Matrix: Flower

Date Reported: 12/1/2023

Prep Analyst:

Client License:

Analysis Method: 0630322+1 H4 4-21-2022 #1.lcm

Sampling Method: N/A

Reference Method: JCB 2009: HPLC/DAD


Analysis Batch: 11-30-2023 H4 185, 389 Flower

For R&D Purposes Only

Total THC (THCA*0.877+d9-THC)	24.5%
Total CBD (CBDA*0.877+CBD)	<LOQ%
Moisture Content	10.9%

Cannabinoid	% Weight	mg/g
CBDVA	<LOQ	<LOQ
CBDV	<LOQ	<LOQ
CBDA*	0.114	1.14
CBGA	3.39	33.9
CBG	0.143	1.43
CBD*	<LOQ	<LOQ
THCV	<LOQ	<LOQ
CBN	<LOQ	<LOQ
d9-THC*	0.182	1.82
d8-THC*	<LOQ	<LOQ
CBC	<LOQ	<LOQ
THCA*	27.7	277.0
Total Cannabinoids	31.53	315.0

*ORELAP Accredited Analyte
Limit Of Quantitation: 0.1%, analyte not measured



- CBDA*
- CBGA
- CBG
- d9-THC*
- THCA*



These test results may not be altered or reproduced except in full without the permission of . These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152 Report generated by Routine_Potency_Rev13_8-1-2023

[Signature]
PhD
Lab Director

Quality Control Results

Analyst: Jeff A.

Analysis Batch: 11-30-2023 H4 185, 389 Flower

	Duplicate RPD		LCS % Recovery		Method Blank	
	H-0-2079-b	Limit	C-FL-11302023	Limits	C-FB-112823	Limit
CBDA	<LOQ%	30%	97.9%	90-110%	<LOQ/2	LOQ/2
CBD	<LOQ%	30%	102.0%	90-110%	<LOQ/2	LOQ/2
d9-THC	11.4%	30%	102.0%	90-110%	<LOQ/2	LOQ/2
d8-THC	<LOQ%	30%	103.0%	90-110%	<LOQ/2	LOQ/2
THCA	3.61%	10%	92.9%	90-110%	<LOQ/2	LOQ/2

RPD: Relative Percent Difference between unknown sample and its duplicate

LCS: Laboratory Control Sample with known concentration

Case Comments: There were no divergences from ordinary Quality Control procedures or SOPs.

MADRONA HARVEST



These test results may not be altered or reproduced except in full without the permission of . These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152 Report generated by Routine_Potency_Rev13_8-1-2023

Jeff A.

 Lab Director, PhD