

Potency Results

Sample Name: *Hemp A - Ice Cream Cake - J*

Client:

Client Batch ID:

Sample ID: rC-H-185-D1804

Date Sampled: 11/1/2023

Matrix: Flower

Date Reported: 11/3/2023

Prep Analyst:

Client License:

Analysis Method: 0630322+1 H4 4-21-2022 #1.lcm

Sampling Method: N/A


Reference Method: JCB 2009: HPLC/DAD

For R&D Purposes Only

Analysis Batch: 11-2-2023 H4 185 Flower

Total THC (THCA*0.877+d9-THC)	26.6%
Total CBD (CBDA*0.877+CBD)	<LOQ%
Moisture Content	10.2%

Cannabinoid	% Weight	mg/g
CBDVA	<LOQ	<LOQ
CBDV	<LOQ	<LOQ
CBDA*	0.106	1.06
CBGA	2.1	21.0
CBG	<LOQ	<LOQ
CBD*	<LOQ	<LOQ
THCV	<LOQ	<LOQ
CBN	<LOQ	<LOQ
d9-THC*	0.121	1.21
d8-THC*	<LOQ	<LOQ
CBC	<LOQ	<LOQ
THCA*	30.2	302.0
Total Cannabinoids	32.53	325.0



- CBDA*
- THCA*
- CBGA
- d9-THC*

*ORELAP Accredited Analyte
Limit Of Quantitation: 0.1%, analyte not measured



These test results may not be altered or reproduced except in full without the permission of . These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152 Report generated by Routine_Potency_Rev13_8-1-2023

[Signature]
PhD
Lab Director

Quality Control Results

Analyst:

Analysis Batch: 11-2-2023 H4 185 Flower

	Duplicate RPD		LCS % Recovery		Method Blank	
	H-0-D1790-b	Limit	C-FL-110223	Limits	C-FB-110223	Limit
CBDA	<LOQ%	30%	97.9%	90-110%	<LOQ/2	LOQ/2
CBD	<LOQ%	30%	102.0%	90-110%	<LOQ/2	LOQ/2
d9-THC	4.75%	30%	102.0%	90-110%	<LOQ/2	LOQ/2
d8-THC	<LOQ%	30%	103.0%	90-110%	<LOQ/2	LOQ/2
THCA	7.04%	10%	92.9%	90-110%	<LOQ/2	LOQ/2

RPD: Relative Percent Difference between unknown sample and its duplicate
 LCS: Laboratory Control Sample with known concentration
 Case Comments: There were no divergences from ordinary Quality Control procedures or SOPs.



These test results may not be altered or reproduced except in full without the permission of . These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152 Report generated by Routine_Potency_Rev13_8-1-2023

[Signature]
 , PhD
 Lab Director